

BLANK PAGE



Indian Standard


SPECIFICATION FOR FORCEPS, PERITONSILLAR

UDC 615·472·41:616·211—089



© Copyright 1968

INDIAN STANDARDS INSTITUTION MANAK BHAVAN, 9 BAHADUR SHAH ZAFAR MARG NEW DELHI 1



Indian Standard SPECIFICATION FOR FORCEPS, PERITONSILLAR

Surgical Instruments Sectional Committee, CPDC 11

Chairman

COL R. D. AYYAR

Memhers

CHIEF SALES MANAGER
WORKS MANAGER (Alternate)
SHRI R. D. GUPTA

DR (KUMARI) S. D. GUPTA

BRIG IQBAL CHAND

SHRI G. B. JAKHETIA SHRI V. KRISHNAMOORTHY SHRI SANTI KUMAR MUKERJI SHRI M. G. PANDIT DR A. RAMDAS

Shri Mahesh M. Shah Maj Sohan Singh

SHRIK, L. TALWAR

SHRI S. SEN GUPTA

SHRI A. N. VIG SHRI R. N. VIG (Alternate)

SHRI A. B. RAO,

Director (Consr Prod.) (Secretary)

Safdarjang Hospital, Central Government Health Scheme (Ministry of Health & Family Planning), New Delhi

Indian Drugs & Pharmaceuticals Ltd, New Delhi

Representing

National Metallurgical Laboratory (CSIR), Jamshedpur

Lady Hardinge Medical College & Hospital, New Delhi

Directorate General of Armed Forces Medical Services (Ministry of Defence)

Development Commissioner, Small Scale Industries Directorate General of Technical Development H. Mukerji & Banerjee Surgical Pvt Ltd, Calcutta Ministry of Health & Family Planning

Ministry of Railways

Baruipur Surgical Instrument Centre (Directorate of Industries, Government of West Bengal)

Shah Medical & Surgical Co Ltd, Baroda Directorate General of Inspection (Ministry of

Defence)
Chief Inspectorate of General Stores [Ministry of Defence (DGI)], Kanpur

ISCOMED Industries Private Ltd, Bombay

Director General, ISI (Ex-officio Member)

Precision Type Surgical Instruments Subcommittee, CPDC 11:5

Convener

DR J. L. GUPTA

Safdarjang Hospital, Central Government Health Scheme (Ministry of Health & Family Planning), New Delhi

Members

Dr Jacob Chandi Dr G. S. Gopalan Christian Medical College, Vellore Jaya Surgicals, Madras

SHRI G. JAYA PRASAD (Alternate)

(Continued on page 2)

INDIAN STANDARDS INSTITUTION MANAK BHAVAN, 9 BAHADUR SHAH ZAFAR MARG NEW DELHI 1

IS: 4504 - 1968

(Continued from page 1)

Members

Dr Gopinath Brig Iqbal Chand

Shri Santi Kumar Mukerji Shri H. C. Mukherjea

Shri J. Prabhakar Dr T. T. Ramalingam

DR P. B. RAO

Representing

All India Institute of Medical Sciences, New Delhi
Directorate General of Armed Forces Medical
Services (Ministry of Defence)
H. Mukerji & Banerjee Surgical Pvt Ltd, Calcutta
Chief Inspectorate of General Stores [Ministry of
Defence (DGI)], Kanpur
Indian Drugs & Pharmaceuticals Ltd, Madras
In personal capacity (4-A Kilpauk Garden Road,
Kilpauk, Madras)
Jawaharlal Institute of Post-Graduate Medical Edu-

Jawaharlal Institute of Post-Graduate Medical Education & Research (Ministry of Health & Family Planning), Pondicherry

Indian Standard SPECIFICATION FOR FORCEPS, PERITONSILLAR

0. FOREWORD

- 0.1 This Indian Standard was adopted by the Indian Standards Institution on 21 February 1968, after the draft finalized by the Surgical Instruments Sectional Committee had been approved by the Consumer Products Division Council.
- 0.2 The formulation of Indian Standards on surgical instruments has been taken up on the recommendations of the Advisory Committee for Development of Surgical Instruments, Equipment and Appliances, Government of India.
- **0.3** This standard is one of a series of Indian Standards on surgical instruments. Other standards published so far in the series appear at P 6.
- 0.4 For the purpose of deciding whether a particular requirement of this standard is complied with, the final value, observed or calculated, expressing the result of a test or analysis, shall be rounded off in accordance with IS: 2-1960*. The number of significant places retained in the rounded off value should be the same as that of the specified value in this standard.

1. SCOPE

1.1 This standard lays down the requirements for peritonsillar forceps with bulbous, regular and trocar points.

2. MATERIAL


- 2.1 The forceps shall be made of stainless steel conforming to Designation 22Cr13 or 30Cr13 of Schedule V of IS: 1570-1961†.
- 2.2 The screw shall be of stainless steel of the Designations prescribed in 2.1 or 20Cr18Ni2 of Schedule V of IS: 1570-1961†.

3. SHAPE AND DIMENSIONS

3.1 The general shape and dimensions of the forceps shall be as given in Fig. 1.

^{*}Rules for rounding off numerical values (revised).

[†]Schedules for wrought steels for general engineering purposes.



*See IS: 3642-1966 'General requirements for surgical instruments'.

All dimensions in millimetres.

Fig. 1 Forceps, Peritonsillar

3.2 Joints — The joints shall either be box type or screw recessed type in accordance with 6 of Section 2 of IS: 3642-1966*.

4. WORKMANSHIP AND FINISH

4.1 The forceps shall be of a symmetrical and balanced construction and shall open and close with even movement without jerks. The triangular tips with regular as well as trocar points shall be sharp on the point and edges semi-sharp. The tips shall meet accurately as to match the serrations in the bulbous point and to give a sharp point in the others. The arms of the forceps shall be smooth and free from surface defects like scales, burrs, pits, cracks and seams. All sharp edges, except those of the tip, shall be rounded smooth. They shall be passivated and polished bright. When the finger loops are pressed on a smooth horizontal surface, the jaws of the forceps shall be horizontal.

5. HEAT TREATMENT

5.1 The entire point of the forceps shall be suitably hardened (see also Fig. 1) and tempered to give a uniform hardness of 430 to 490 HV.

6. TEST

- 6.1 Corrosion Resistance The forceps shall be tested for corrosion resistance as given in 6.1.1.
- 6.1.1 Copper Sulphate Test The sample shall be scrubbed with soap and warm water, rinsed in hot water, followed by a dip in ethyl alcohol (95 percent) and dried. The sample shall be completely immersed in copper sulphate solution for six minutes and shall then be washed off with fresh water or with wet cotton wool. The copper sulphate solution shall be made up as follows:

Copper sulphate (CuSO ₄ .5H ₂ O)	4.0 g
Sulphuric acid (H ₂ SO ₄) (sp gr 1.84)	10·0 g
Water (H.O)	90·0 ml

There shall be no red stains or spots on the sample after the test, but the polished surface getting dull may be permitted.

7. MARKING

7.1 Each forceps shall be clearly and indelibly marked with the manufacturer's name, initials or trade-mark and the words 'STAINLESS STEEL'.

^{*}General requirements for surgical instruments.

IS: 4504 - 1968

7.1.1 The forceps may also be marked with the ISI Certification Mark.

Note — The use of the ISI Certification Mark is governed by the provisions of the Indian Standards Institution (Certification Marks) Act, and the Rules and Regulations made thereunder. Presence of this mark on products covered by an Indian Standard conveys the assurance that they have been produced to comply with the requirements of that standard, under a well-defined system of inspection, testing and quality control during production. This system, which is devised and supervised by ISI and operated by the producer, has the further safeguard that the products as actually marketed are continuously checked by ISI for conformity to the standard. Details of conditions, under which a licence for the use of the ISI Certification Mark may be granted to manufacturers or processors, may be obtained from the Indian Standards Institution.

8. PACKING

8.1 The forceps shall be wrapped in moisture-proof paper or polyethylene bag. The tips shall be suitably protected. They shall be packed in a manner that coming into contact with each other is avoided.

INDIAN STANDARDS

ON

Surgical Instruments

IS:				Rs
3318-1965	General requirements for surgical scalpels and knives			1.00
3319-1965	Surgical scalpels with detachable blades (Bard Parker	type)		3.00
3320-1965	Surgical scalpels		•••	1.00
3642-1966	General requirements for surgical instruments		•••	2.50
3643-1966	Dissecting forceps (serrated and toothed)		•••	2.00
3644-1966	Artery forceps (Halsted Mosquite pattern)		•••	1.50
3645-1966	Artery forceps (Spencer Wells pattern)	•••	•••	1.50
3758-1966	Hooks, aural	•••	•••	1.20
3759-1966	Blades, skin grafting	•••	•••	1.00
3760-1966	Probe, aural, Jobson Horne's pattern		•••	1.00
3767-1966	Hook, Skin, Gillies pattern		•••	1.00
3787-1966	Gag, mouth, Cleft palate	•••	•••	1.50
3788-1966	Specula, aural	•••	•••	1.50
3799-1966	Forceps, aural and nasal, crocodile action	•••	•••	2.00
3807-1966	Forceps, dressing, aural and nasal	•••	•••	1.50
4085-1967	Knives, amputating	•••	•••	2.00
4089-1967	Forceps, clip removing Michel's	•••	•••	2 ·50
4094-1967	Forceps, sterilizer (Cheatle's)	•••	•••	2 ·50
4244-1967	Retractors, skin, double ended	•••	•••	2.00
4245-1967	Needle holder (Gillies) combined with scissors, for	plastic s	urgery	2 ·50
4275-1967	Scissors, Kilner's for plastic surgery	•••	•••	2 ·50
4281-1967	Scissors, Mc Indoe's, for plastic surgery	•••	•••	3 ·50
4282-1967	Forceps, dissecting, plastic surgery	•••	•••	3.50
4487-1968	Forceps, tonsil artery (Birkett's pattern)	•••	•••	
4488-1968	Forceps, tonsil holding (Muck's and White's pattern)	•••	•••	
4489-1968	Forceps, nasal turbinate, Luc's	•••	•••	
4490-1968	Forceps, dilating, Troussceu's	•••	•••	
4513-1968	Scissors, surgical dressing and stitch	•••	•••	

PUBLICATIONS OF INDIAN STANDARDS INSTITUTION

INDIAN STANDARDS

About 4 500 Indian Standards, broadly classified under the following main heads, have been issued so far:

Agriculture & Food Chemical Civil Engineering Consumer Products

Electrotechnical
Mechanical Engineering
Structural & Metals
Textile

Of these, the standards belonging to the Consumer Products Group, at present, fall under the following categories:

Colr and Coir Products

Domestic Hardware

Medical Instruments and Hospital Equipment

Oil and Gas
Appliances

Sports Goods and Gymnasium Equipment

Umbrellas

Safety Matches

Utensils and Cutlery

Writing Aids

OTHER PUBLICATIONS

Burning

ISI Bulletin (Published Every Month	N3			
Single Conv	•			2•00
	•••	•••	•••	15.00
Annual Subscription	• • • •	•••	•••	,
Annual Reports (from 1948-49 Onw	ards)	•••	•••	2.00 to 3.00 each
Handbook of ISI Publications, 1967	•••	•••	•••	5.00

Available from

INDIAN STANDARDS INSTITUTION

Headquarters

Manak Bhavan, 9 Bahadur Shah Zafar Marg, New Delhi I

Telephones 27 36 11 - 20

Telegrams Manaksanstha

Brunch Offices Telegrams Manaksanstha

534 Sardar Vallabhbhal Patel Road	Bombay 7	Telephone	35 69 44
5 Chowringhee Approach	Calcutta 13	,,	23-18 23
Industrial Estate Administrative	Hyderabad 18	**	3 92 29
Building, Sanatnagar			
117/418 B Sarvodaya Nagar	Kanpur	9,	3 76 95
54 General Patters Road	Madras 2	••	8 72 78

AMENDMENT NO. 1 JULY 1975

TO

IS:4504-1968 SPECIFICATION FOR FORCEPS, PERITONSILLAR

Alterations

(Page 5, clauses 6.1 and 6.1.1) - Substitute the following for the existing clauses:

'6.1 Corrosion Resistance - The instrument shall satisfy the boiling and autoclaving test as mentioned in IS:7531-1975.'

[Page 5, foot-note with asterisk(4)] - Substitute the following after foot-note with asterisk(*):

(CPDC 11)

Method for boiling and autoclaving test
for corrosion resistance of stainless
steel surgical instruments.*

Reprography Unit, BIS, New Delhi, India